

ELECTRICAL SAFETY RECALL

3 Phase App/Ext/RCD

AERO2032 RCD Lead Adaptor Box
Sold between 2014 and 2020



IDENTIFICATION: The unit is designed to conduct the required tests for 3 phase equipment, RCDs and Extension Leads using a single-phase appliance tester.
Part number: 3PHEXTRCD

DEFECT: When the device is plugged into 3 phase power, the exposed 3 phase lead can become live if not plugged into unused 3 phase socket.

HAZARD: There is a potential electrical shock hazard

WHAT TO DO: You need to cease use immediately and return the unit to the place of purchase for a modification to the product. All costs associated with sending and then receiving it will be paid for.

CONTACT DETAILS: Contact Appliance Testing Supplies on 1300 656 938 between Mon-Fri 8:30am-5:00pm, via email at sales@appliance-testingsupplies.com.au or go to www.portableappliance-testers.com.au

For further information

Please direct all calls and any queries concerning this recall to Appliance Testing Supplies on 1300 656 938.

See www.productsafety.gov.au for Australian product recall information